



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/726,558	12/01/2000	Ryo Ozawa	P20020	1830

7055 7590 08/08/2006

GREENBLUM & BERNSTEIN, P.L.C.
1950 ROLAND CLARKE PLACE
RESTON, VA 20191

EXAMINER

CZEKAJ, DAVID J

ART UNIT	PAPER NUMBER
----------	--------------

2621

DATE MAILED: 08/08/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

09/726,558

Applicant(s)

OZAWA, RYO

Examiner

Dave Czekaj

Art Unit

2621

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 14 June 2006.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-12 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-12 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Response to Arguments

1. Applicant's arguments with respect to the rejection(s) of claim(s) 1-12 have been fully considered and are persuasive. Therefore, the rejection has been withdrawn. However, upon further consideration, a new ground(s) of rejection is made as set forth below.

Claim Rejections - 35 USC § 103

2. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

3. Claims 1-3, 5-9, and 11-12 rejected under 35 U.S.C. 103(a) as being unpatentable over Kanno et al. (5583566), (hereinafter referred to as "Kanno") in view of Nishikori et al. (5627584), (hereinafter referred to as "Nishikori").

Regarding claims 1 and 7, Kanno discloses an apparatus for interfacing a medical instrument wherein no cable for transmitting information is required (Kanno: column 4, lines 63-65). This apparatus comprises an "endoscope having a solid state image sensor provided at a distal end, an image signal processing unit that produces a video signal based on the image signals, and a monitor for reproducing and displaying the images" (Kanno: column 7, lines 43-54, wherein the solid state image sensor is the CCD which is located on the tip or distal end and the signal processing unit is the conversion of the input signal to a

video signal). The system further comprises a "scene changing system that changes a scene on the monitor between an endoscope image display scene and a patient data list display scene" (Kanno: figure 32, wherein the user has the ability to change the scene between the endoscope image display or endoscope inspection and the patient data list or patient data management), "storage system that stores patient data forming a patient data list" (Kanno: figure 30A, column 22, lines 49-55, wherein the storage device is the hard disc), "a selection system that selects individual patient data from the patient data list displayed on the monitor" (Kanno: figure 32, column 23, lines 53-67 – column 24, lines 1-67, wherein the selection system is the program displayed on the screen in figure 32), and a "display control system that displays the individual patient data together with the endoscope image on the monitor when the scene is changed from the patient list to the endoscope image display" (Kanno: figures 24 and 33, wherein the endoscope images are displayed in box 203a). However, this apparatus lacks displaying the patient data list on a monitor. Nishikori teaches that prior art endoscope systems make the operating procedure more complex (Nishikori: column 1, lines 48-52). To help alleviate this problem, Nishikori discloses "a patient data list which is displayed on the monitor" (Nishikori: figures 15D, 15F, 15I, and 15J). Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was made to take the apparatus disclosed by Kanno and add the patient list display taught by Nishikori in order to obtain an apparatus that makes operating procedures less complex.

Regarding claims 2 and 8, Kanno discloses an “editing system that edits the patient data forming the patient data list” (Kanno: column 25, lines 21-25, wherein the editing system is the patient data management).

Regarding claims 3 and 9, Nishikori discloses “the production of the video signal is performed by the image signal processing unit such that as much patient information as possible is included in the patient data list to be displayed on the monitor when the scene is changed from the endoscope image display to the patient data list display” (Nishikori: figure 15D, wherein the endoscope image display is invoked by pressing the CV-100 button, figure 15I, wherein the screen is shown to occupy the entire screen to display as much information as possible).

Regarding claims 5 and 11, Kanno discloses that the selection system further includes an “indicator system that visually indicates patient data to be selected from the patient data list” (Kanno: column 23, lines 47-55, wherein the visual indicator is the mouse), “manual operation system that controls the indication of the patient data to be selected from the list” (Kanno: figure 32, column 23, lines 53-67 – column 24, lines 1-67, wherein the operating system is the program that runs the menu displayed on the screen in figure 32) and a “manual settlement system that manually settles the indication of the patient data to be selected from the patient data list” (column 23, lines 47-55, wherein the settlement system is the mouse in that the mouse “click” manually settles or selects the appropriate data).

Regarding claims 6 and 12, Kanno discloses an “editing system that edits the patient data forming the patient list” (Kanno: column 25, lines 21-25, wherein the editing system is the patient data management), and a “determination system that determines whether the editing of the patient data is performed by an editing system after the activation of the manual settlement system, the editing of the patient data being settled by an activation of the manual settlement system when the performance of the editing of the patient data is confirmed by the determination system” (Kanno: column 23, lines 47-55, wherein the settlement system is the mouse in that the mouse “click” manually settles or selects the appropriate data. The data will not be edited until the selection is “clicked” or confirmed by the system).

4. Claims 4 and 10 are rejected under 35 U.S.C. 103(a) as being unpatentable over Kanno et al. (5583566), (hereinafter referred to as “Kanno”) in view of Nishikori et al. (5627584), (hereinafter referred to as “Nishikori”) in further view of Ozawa et al. (6154248), (hereinafter referred to as “Ozawa”).

Regarding claims 4 and 10, note the examiners rejection for claims 1 and 7, and in addition, although one of ordinary skill would realize that Nishikori’s apparatus would utilize clock signals to transfer data, Nishikori fails to disclose the specifics of the clock signals as claimed. Ozawa teaches that connecting an endoscope to peripheral equipment based on one frequency can lead to a deteriorated quality of the reproduced color image (Ozawa: column 2, lines 66-67 – column 3, lines 1-36). To fix this problem, Ozawa discloses an apparatus

Art Unit: 2621

that outputs digital images based on a second series of clock pulses having a frequency different from that of the first series of clock pulses (Ozawa: column 3, lines 39-49). Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was made to take the apparatus disclosed by Nishikori, add the display system taught by Kanno, and add the multiple clock pulses with different frequencies taught by Ozawa in order to obtain an apparatus that produces superior video quality. One would be further motivated since Nishikori is silent on how to do so.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Dave Czekaj whose telephone number is (571) 272-7327. The examiner can normally be reached on Monday - Friday 9 hours.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Mehrdad Dastouri can be reached on (571) 272-7418. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 2621

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

DJC

Mehrdad Dastouri
MEHRDAD DASTOURI
SUPERVISORY PATENT EXAMINER
TC 2600